

# Analysis of Low Concentrations of Perchlorate in Drinking Water and Ground Water by Ion Chromatography

## **INTRODUCTION**

Perchlorate (as ammonium perchlorate), which is widely used in solid rocket propellants, has recently been found in drinking water wells in areas where aerospace materials and munitions have been manufactured and tested.<sup>1</sup> Perchlorate is a health concern, as it interferes with the production of thyroid hormones. Current data suggest that an exposure level range of 4 to 18 µg/L (ppb) is acceptable.<sup>2</sup> Although perchlorate is not yet regulated in the U.S. under the Federal Safe Drinking Water Act, the State of California requires remedial action for drinking water sources containing greater than 18 µg/L of perchlorate.

This Application Note details a new method developed to quantify low levels of perchlorate. A large loop injection (1000 µL) is used with an IonPac® AS11 column and suppressed conductivity detection to quantify perchlorate in drinking water down to approximately 2.5 µg/L.

## **EQUIPMENT**

Dionex DX-500 Ion Chromatography system consisting of:

- GP40 Gradient Pump
- CD20 Conductivity Detector
- AS40 Automated Sampler
- LC20 Chromatography Enclosure with a rear-loading valve

4-L Plastic bottle assemblies (two for external water mode)

PeakNet Chromatography Workstation

## **REAGENTS AND STANDARDS**

Deionized water (DI H<sub>2</sub>O), Type I reagent grade, 18 MΩ-cm resistance or better

Sodium hydroxide, 50% (w/w) aqueous solution (Fisher Scientific or other)

Sodium perchlorate, 99% ACS reagent grade or better (Aldrich or other)

Potassium sulfate, 1000 mg/L aqueous solution (Ultra Scientific or other)

## **CONDITIONS**

Columns:	IonPac AS11 Analytical, 4 x 250 mm (P/N 44076) IonPac AG11 Guard, 4 x 50 mm (P/N 44078)
Eluent:	100 mM Sodium hydroxide
Run Time:	12 min
Flow Rate:	1.0 mL/min
Sample Volume:	1000 µL
Detection:	Suppressed conductivity, ASRS® (4 mm), AutoSuppression® external water mode
System	
Backpressure:	600–900 psi (3.95–5.93 MPa)
Background	
Conductance:	2–5 µS

Now sold under the  
Thermo Scientific brand

**Thermo**  
SCIENTIFIC

## PREPARATION OF SOLUTIONS AND REAGENTS

### Standard Solution

*Stock perchlorate standard solution (1000 mg/L)*

Dissolve 1.231 g of sodium perchlorate in 1000 mL of deionized water to prepare a 1000 mg/L standard. Standard is stable for at least one month when stored at 4 °C.

### Working Standard Solutions

Dilute 1000 mg/L standard solution as required with deionized water to prepare the appropriate working standards.

### Eluent Solution

*100.0 mM Sodium hydroxide*

Weigh 992.0 g of deionized water into an eluent bottle. Degas water for approximately 5 minutes. Carefully add 8.0 g of 50% sodium hydroxide directly to the bottle. Mix then quickly transfer the eluent bottle to the instrument and pressurize the bottle with helium at 8 psi (0.055 MPa).

## RESULTS AND DISCUSSION

For the best performance at low-ppb levels, it is critical that baseline noise be kept to a minimum. To minimize baseline noise, it is necessary to use the ASRS in external water mode rather than the recycle mode. An equilibrated system will produce a background conductance between 2–5  $\mu$ S. Peak-to-peak noise is typically 10 nS and system backpressure is 600–900 psi (3.95–5.93 MPa). A system blank is determined by using deionized water as a sample. This blank establishes the baseline and confirms the lack of contamination in the system. The linear concentration range was determined to ensure accurate quantification of perchlorate in the 2.5–100  $\mu$ g/L range. Figure 1 shows the results of a linearity study.




Figure 1 Perchlorate calibration

This plot demonstrates that calibration of perchlorate is linear in the low-ppb range. Figure 2 shows a typical chromatogram of a 20  $\mu$ g/L perchlorate standard. To determine the method detection limit (MDL), seven injections of the 2.5  $\mu$ g/L perchlorate standard were made. Table 1 shows the results of a method detection limit study. The 1000  $\mu$ L injection is large enough to achieve the desired detection limit without overloading the column. Note that this method is not intended for use with high (ppm) levels of perchlorate. The calculated MDL equals 254 ng/L (ppt).




Figure 2 20  $\mu$ g/L Perchlorate standard

Table 1 MDL for perchlorate based on a 1000  $\mu$ L injection volume

Injection #	Area counts	Retention time (min)
1	3391	9.48
2	3405	9.57
3	3504	9.50
4	3503	9.45
5	3435	9.47
6	3301	9.52
7	3315	9.43
Average	3408	9.49
SD	81	0.05
RSD	2.38	0.49

MDL=254 ng/L (ppt), MDL=SD\* $t_{s,99}$  where  $t_{s,99}$ =3.14 for n=7

Figures 3 through 5 show chromatograms obtained for 2.5 µg/L perchlorate in three different matrices. Figure 3 shows the chromatogram of 2.5 µg/L perchlorate in deionized water. Figure 4 shows 2.5 µg/L perchlorate in tap water. Note that all other anions present in tap water elute in the void volume and do not interfere with perchlorate determination. Some environmental samples may contain low levels of perchlorate in the presence of a large amount of sulfate. Figure 5 shows the determination of 2.5 µg/L perchlorate in the presence of 700 mg/L sulfate. The high concentration of sulfate does not affect perchlorate recovery or the detection limit.

### SUMMARY

The method outlined in this Application Note allows the determination of low-µg/L (ppb) levels of perchlorate. Linear concentration ranges have been established to accurately quantify perchlorate in drinking water and ground water samples.




Figure 3 2.5 µg/L Perchlorate standard




Figure 4 2.5 µg/L Perchlorate in tap water




Figure 5 2.5 µg/L Perchlorate and 700 mg/L Sulfate

---

## REFERENCES

1. "Perchlorate in California Drinking Water"; California Department of Health Services, September 1997.
2. Correspondence from Joan S. Dollarhide, National Center for Environmental Assessment, Office of Research and Development, to Mike Girrard, Chairman, Perchlorate Study Group, U.S. EPA, 1995.

## LIST OF SUPPLIERS

Aldrich Chemical Company, Inc., 1001 West Saint Paul Avenue, P.O. Box 355, Milwaukee, Wisconsin, 53233, USA. Tel: 1-800-558-9160.

Fisher Scientific, 711 Forbes Ave., Pittsburgh, Pennsylvania, 15219-4785, USA.  
Tel: 1-800-766-7000.

Ultra Scientific, 250 Smith Street, North Kingstown, Rhode Island, 02852, USA. Tel: 401-294-9400.



AutoSuppression, ASRS, and IonPac are registered trademarks of Dionex Corporation.

Printed on recycled and recyclable paper.

---

**Dionex Corporation**  
1228 Titan Way  
P.O. Box 3603  
Sunnyvale, CA  
94088-3603  
(408) 737-0700

**Dionex Corporation**  
Salt Lake City Technical Center  
1515 West 2200 South, Suite A  
Salt Lake City, UT  
84119-1484  
(801) 972-9292

**Dionex U.S. Regional Offices**  
Sunnyvale, CA (408) 737-8522  
Westmont, IL (630) 789-3660  
Houston, TX (281) 847-5652  
Atlanta, GA (770) 432-8100  
Marlton, NJ (856) 596-0600

**Dionex International Subsidiaries**  
*Austria* (01) 616 51 25 *Belgium* (015) 203800 *Canada* (905) 844-9650 *France* 01 39 46 08 40 *Germany* 06126-991-0  
*Italy* (06) 66 51 50 52 *Japan* (06) 6885-1213 *The Netherlands* (0161) 43 43 03 *Switzerland* (062) 205 99 66 *United Kingdom* (01276) 691722  
\* Designed, developed, and manufactured under an NSAI registered ISO 9001 Quality System.  
<http://www.dionex.com>



LPN 0991-02 1.2M 6/00  
© 2000 Dionex Corporation